

ANNEXURE-1 PAGE 1 OF 2 TO REGISTRATION CERTIFICATE

MANUFACTURER'S NAME : EXCEL INSTRUMENTS

**Registered Office: 52, Haridwar Society,
Detroj Road, Kadi- 382 715, Tal: Kadi,
Dist: Mehsana, Gujarat, INDIA
Works: C - 1/1303, Phase 1, G.I.D.C.,
Chhatral - 382 729, Tal: Kalol,
Dist. Gandhinagar, Gujarat, INDIA**

APPLIED DIRECTIVE

**: LOW VOLTAGE DIRECTIVE 2006/95/EC
PRESSURE EQUIPMENT DIRECTIVE 97/23/EC
Scope**



❖ PRESSURE GAUGE :

Category	Nominal Size	Mounting	Vacuum / Pressure Range
Bourdon Type	50mm to 150mm	Bottom Direct, Bottom Surface, Back Direct, Back Panel	0.6 to 1000 Kg/cm ² Vacuum (-)760 mm of hg
Sealed Diaphragm Type	63mm to 150mm	Bottom Direct, Bottom Surface, Back Direct, Back Panel, Remote Type	0.6 to 700 Kg/cm ² Vacuum (-)760 mm of hg
Low Pressure Diaphragm Type	100mm to 150mm	Bottom Direct	250 to 6000 mm of wc (-) 250 To -6000 mm of wc
Schaffer Diaphragm Type	100mm to 150mm	Bottom Direct	1 to 21 Kg/cm ² Vacuum (-)760 mm of hg
Low Pressure Capsule Type	63mm to 150mm	Bottom Direct, Back Direct & Back Panel	160 to 6000 mm of wc (-) 160 To -6000 mm of wc
Differential Pressure Gauge - Bourdon type	150mm	Bottom Direct	0.6 to 10 Kg/cm ²
Differential Pressure Gauge - Diaphragm type	100mm to 150mm	Back Direct & Bottom Direct, Wall Mounting, Pipe Mounting	1 to 21 Kg/cm ² 250 to 6000 mm of wc
Differential Pressure Gauge - Magnetic Piston type	63mm to 150mm	Bottom & In-Line	0.6 to 10 Kg/cm ²
Electrical Contact Type Pressure Gauge	100mm to 150mm	Bottom Direct, Bottom Surface, Back Direct, Back Panel	0.6 to 1000 Kg/cm ² Vacuum (-)760 mm of hg
PP Case Solid Front Pressure Gauge	115mm	Bottom & Wall Mounting	0.6 to 1000 Kg/cm ² Vacuum (-)760 mm of hg

**Authorised Signatory
Managing Director**


Certificate No : GUJ / C - 3133

Date : 20 October 2015

Valid Until : 19 October 2018

ANNEXURE-1 PAGE 2 OF 2 TO REGISTRATION CERTIFICATE

MANUFACTURER'S NAME : EXCEL INSTRUMENTS
Registered Office: 52, Haridwar Society,
Detroj Road, Kadi- 382 715, Tal: Kadi,
Dist: Mehsana, Gujarat, INDIA
Works: C - 1/1303, Phase 1, G.I.D.C.,
Chhatral - 382 729, Tal: Kalol,
Dist. Gandhinagar, Gujarat, INDIA



APPLIED DIRECTIVE : LOW VOLTAGE DIRECTIVE 2006/95/EC
PRESSURE EQUIPMENT DIRECTIVE 97/23/EC

Scope

❖ <u>TEMPERATURE GAUGE :</u>			
Gas Filled	100mm to 150mm	Bottom Direct, Bottom Surface, Back Direct, Back Panel	(-) 30 To 650 °C
Mercury Filled	100mm to 150mm	Bottom Direct, Bottom Surface, Back Direct, Back Panel	(-) 30 To 650 °C
Bi - Metal	63mm to 150mm	Bottom Direct, Back Direct	(-) 30 To 500 °C
Gas Filled Temperature Gauge With Electrical Contact	100mm to 150mm	Bottom Direct, Bottom Surface, Back Direct, Back Panel	(-) 30 To 650 °C
Digital	100mm	Bottom Direct, Bottom Surface, Back Direct, Back Panel	(-) 100 To 450 °C (-) 50 to 1000 °C

Authorised Signatory
Managing Director

Certificate No : GUJ / C - 3133
Date : 20 October 2015
Valid Until : 19 October 2018